



EU Quality Management System Certificate (IVDR)

Pursuant to Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices,
Annex IX Chapters I and III (Class C and B Devices excluding self-/near-patient-testing and
Companion Diagnostics)

No. V12 113961 0002 Rev. 00

Manufacturer: **Alba Bioscience Limited**
5 James Hamilton Way
Milton Bridge, Penicuik, EH26 0BF
UNITED KINGDOM

SRN Manufacturer: GB-MF-000029300

Authorized Representative: Emergo Europe B.V.
Westervoortsedijk 60, 6827 AT Arnhem, THE NETHERLANDS

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (8) of the Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices. Details on devices covered by the quality management system are described on the following page(s).

The Report referenced below summarizes the result of the assessment and includes reference to relevant CS, harmonized standards, audit and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result.

The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment includes an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:V12 113961 0002 Rev. 00](http://www.tuvsud.com/ps-cert?q=cert:V12_113961_0002_Rev.00)

Report No.: 713230840

Valid from: 2023-04-21

Valid until: 2028-04-20

Issue date: 2023-04-21

Marta Carnielli
Head of Notified Body IVD



EU Quality Management System Certificate (IVDR)

Pursuant to Regulation (EU) 2017/746 on in Vitro Diagnostic Medical Devices, Annex IX Chapters I and III (Class C and B Devices excluding self-/near-patient-testing and Companion Diagnostics)

No. V12 113961 0002 Rev. 00

Classification: Class C
Device Group: W0103 - HAEMATOLOGY / HAEMOSTASIS / IMMUNOHAEMATOLOGY / HISTOLOGY / CYTOLOGY
IVP Code: IVP 3001 - In vitro diagnostic devices which require knowledge regarding agglutination tests
Intended Purpose: IVR 0106 - Other devices intended to be used for blood grouping

The validity of this certificate depends on conditions and/or is limited to the following: -none-

Revision History:

Rev.	Dated	Report	Description
00	2023-04-21	713230840	Initial issuance