



# ALBAcyte<sup>®</sup> RhD Negative CAT Reagent Red Cells for ANTIBODY SCREENING

REF Z466



1434

## INTERPRETATION OF LABEL SYMBOLS



Batch code



Use by (YYYY-MM-DD)



Storage temperature limitation (2-8 °C)



*In vitro* diagnostic medical device



Product code



Consult instructions for use

[www.quotientbd.com](http://www.quotientbd.com)



Manufacturer

## INTENDED PURPOSE

The reagent red cells are intended for the detection of irregular red cell antibodies in blood samples from patients who have received prophylactic anti-D.

## INTRODUCTION

Screening patient samples for irregular blood group antibodies is an essential component of compatibility, ante-natal and donor testing protocols. Requirements for antibody screening of patient and donor samples differ and it is acknowledged that implementation of modern blood bank practices demands the use of a sensitive antibody screening

procedure. In this respect the quality of reagent red cells is of paramount importance.

For antibody screening of patient samples, reagent red cells should not be pooled and should display homozygous expression of a range of blood group antigens.

## REAGENT DESCRIPTION

These reagent red cells were prepared from blood donated by three Group O donors and are available as a suspension of washed red cells in a preservative solution ready for use in Column Agglutination Technology (CAT) test systems. The preservative solution contains antibiotics and has been specially formulated to preserve red cell integrity and antigenicity.

The presumptive Rh genotypes of these reagent red cells is r'r, r'r and rr. The full antigenic profile of the individual donations is shown on the enclosed antigen profile. One or more of these red cells may have been held in frozen storage until required. This reagent complies with the requirements of Directive 98/79/EC on *in vitro* Diagnostic Medical Devices and the recommendations contained in the Guidelines for Blood Transfusion Services in the United Kingdom.

## PRECAUTIONS FOR USE AND DISPOSAL

Source material from which this product is derived was found non reactive for HBsAg, Anti-HIV 1/2 and Anti-HCV. **No known test method can offer assurances that products derived from human blood will not transmit infectious disease; therefore, appropriate care should be taken in the use and disposal of this product.**

This reagent is for *in vitro* professional use only.

## STORAGE CONDITIONS

The reagent should be stored at 2-8 °C. Do not freeze. Do not use if obviously discoloured or haemolysed. Do not use beyond the notified expiry date.

## SPECIMEN COLLECTION AND PREPARATION

Specimens should be collected by aseptic technique. The specimen should be tested as soon as possible after collection. If testing is delayed, the specimen should be stored at 2-8 °C for a maximum of 48 hours. Blood specimens exhibiting gross haemolysis or contamination should not be used.

## TEST PROCEDURES

### General Information

CAT and LIS test procedures offer the potential for increased test sensitivity with decreased incubation time and are, therefore, well suited to emergency and routine blood bank situations.

This reagent has been standardised and validated for use by column agglutination techniques, therefore its suitability for use in other techniques cannot be guaranteed. Users are advised to carefully confirm reagent suitability before using alternative techniques.

**For use on the Ortho BioVue system the vial must be aligned such that the barcode printed on the left edge of the vial label (below the red dot) may be read by the system. For use on the BioRad system the vial must be aligned such that the barcode printed on the right edge of the vial label (below the green dot) may be read by the system. If the incorrect barcode is scanned due to mis-positioning of the vial, the information will not be recognised.**

## Materials provided

- ALBAcyte<sup>®</sup> RhD Negative CAT Reagent Red Blood Cells for Antibody Screening (Ref: Z466)

## PERFORMANCE LIMITATIONS

The reaction characteristics of blood group antibodies vary according to their specificity and therefore no single technique will detect all blood group antibodies.

The reactivity of the product may decrease during the dating period and, therefore, the product should not be used after the expiration date. The rate at which the antigen reactivity (e.g. agglutinability) is lost is partially dependent upon individual donor characteristics that are neither controlled nor predicted by the manufacturer.

Antibodies specific for low incidence antigens not present on the test cells will not be detected.

Unwanted positive reactions may occur as a result of red cell donations used in the manufacture of this product, which are untested for HLA Class I (Bg) antigens or associated antigens. False positive or false negative results can occur due to contamination of test materials, improper reaction temperature, improper storage of materials, omission of test reagents and certain disease states.

## SPECIFIC PERFORMANCE CHARACTERISTICS

The reagent red cells have been shown to have a negative direct antiglobulin test, indicating that no human IgG or C3 complement components are detectable on the cell surface.

## DATE OF ISSUE

2023-07

For further information or advice please contact your local distributor.



Emergo Europe B.V.  
Westervoortdijk 60  
6827 AT, Arnhem  
The Netherlands



Alba Bioscience Limited  
James Hamilton Way  
Penicuik  
EH26 0BF, UK

Tel No: +44 (0) 131 357 3333  
Fax No: +44 (0) 131 445 7125  
E-Mail: [customer.serviceEU@quotientbd.com](mailto:customer.serviceEU@quotientbd.com)

## INSTRUCTIONS FOR USE

Instructions for Use are available on Quotient website at [www.quotientbd.com](http://www.quotientbd.com) or can be requested from your Local Distributor, by providing the relevant Product Code stated on product labels and Instructions for Use supplied with the product.

## FEUILLET TECHNIQUE

Le feuillet technique est disponible sur le site Web de Quotient à l'adresse [www.quotientbd.com](http://www.quotientbd.com). Vous pouvez également le demander à votre distributeur local en renseignant le code produit concerné, qui est mentionné sur l'étiquette du produit et dans le feuillet technique fourni avec ce dernier.

## BRUKSANVISNING

Du finner bruksanvisning på Quotient-nettstedet på [www.quotientbd.com](http://www.quotientbd.com), eller du kan be om den hos din lokale forhandler ved å oppgi den aktuelle produktkoden som er angitt i produktmerkingen og i bruksanvisningen som fulgte med produktet.

## KÄYTTÖOHJEET

Käyttöohjeet ovat saatavilla Quotientin sivustossa osoitteessa [www.quotientbd.com](http://www.quotientbd.com), tai ne voi pyytää paikalliset jälleenmyyjältä antamalla tuotteen tuotemerkinnöissä ja mukana toimitetuissa käyttöohjeissa ilmoitetun tuotekoodin.

## GEbruiksaanwijzing

De gebruiksaanwijzing is beschikbaar op de website van Quotient via [www.quotientbd.com](http://www.quotientbd.com) of kan worden aangevraagd bij uw plaatselijke distributeur door de relevante productcode op te geven die staat vermeld op de productlabels en de gebruiksaanwijzing die bij het product is geleverd.

## BRUKSANVISNING

En bruksanvisning är tillgänglig på Quotients webbsida på [www.quotientbd.com](http://www.quotientbd.com) eller kan beställas från din lokala återförsäljare genom att du uppger tillämplig produktkod som anges på produktetiketterna och i den bruksanvisning som levereras med produkten.

## BRUGSANVISNING

Du kan finde en brugsanvisning på Quotients websted på [www.quotientbd.com](http://www.quotientbd.com), eller den kan fås på anmodning fra din lokale distributor ved at oplyse den relevante produktkode på produktmærkatenerne og den brugsanvisning, der fulgte med produktet.